

Eurofins Bactimm B.V.  
Middenkampweg 19  
Nijmegen, 6545 CH  
NL

Client Account Number: A0052261218L  
Eurofins Quote Number: VZB22018067701

**Eurofins Sample Number LV18AB4637-1**

<b>Original Received Date:</b>	25-Sep-2018
<b>Description:</b>	Cellulose Hydrogel uncrosslinked
<b>Lot Number:</b>	35

Analysis	Result	Unit
----------	--------	------

**Cytotoxicity, direct contact test - ISO 10993-5:2009**

RESULT:	Not cytotoxic	----
Applied quantity:	50µL	----
Vehicle:	Supplemented culture medium	----
Positive control:	SDS	----
Negative control:	DPBS	----
Notes:	N/A	----

Addendum #1: Qualitative and quantitative evaluation

Addendum #2: Test for in vitro cytotoxicity

Method: EN ISO 10993-5

Analysis Date: 01-Oct-2018 to 03-Oct-2018

**Contracted Company: Eurofins Biolab Srl (Vimodrone)**

Via Bruno Buozzi, 2, Vimodrone, MI 20090 Italy  
InfoFarma@eurofins.com

Questions about this report should be directed to your project manager or the general email listed above.

	In vitro cytotoxicity ISO10993-5:2009	1-P-PR-TEM-9005812
	Test by direct contact-NRU	Addendum N. 1

ID study // ID sample LV18AB4637-1  
 Test product Cellulose Hydrogel uncrosslinked Batch 35  
 Receiving // Date 25/09/2018  
 Test starting on 01/10/2018 Test finished on 03/10/2018

### QUALITATIVE EVALUATION

#### 1- Qualitative morphological grading of cytotoxicity of

	Contact time:24 hours					
	Replica 1	Replica 2	Replica 3	Replica 4	Replica 5	Replica 6
Vehicol control	0	0	0	0	0	0
Negative control	0	0	0	0	0	0
Positive control	4	4	4	4	4	4
Test product	0	0	0	0	0	0

#### 2- Interpretation of results

Grade	Reactivity	Conditions of all Cultures
0	None	No detectable zone around or under specimen
1	Slight	Some malformed or degenerated cells under specimen
2	Mild	Zone limited to area under specimen
3	Moderate	Zone extending specimen size up to 1,0cm
4	Severe	Zone extending farther than 1,0cm beyond specimen

#### 3-Acceptability criteria

Negative control	
Grade ≤1	VALID

Positive control	
Grade ≥3	VALID

#### 4- Results

	Reactivity grade	
Test product	0	NOT CYTOTOXIC

	In vitro cytotoxicity ISO10993-5:2009	1-P-PR-TEM-9005812
	Test by direct contact-NRU	Addendum N. 1

ID study // ID sample LV18AB4637-1  
 Test product Cellulose Hydrogel uncrosslinked Batch 35  
 Receiving // Date 25/09/2018  
 Test starting on 01/10/2018 Test finished on 03/10/2018

### QUANTITATIVE EVALUATION

#### 1- Optical Density Measurements - OD Value

	Contact time:24 hours					
	Replica 1	Replica 2	Replica 3	Replica 4	Replica 5	Replica 6
Vehicol control	1,315	1,4	1,314	1,209	1,338	1,306
Negative control	1,136	1,269	1,18	1,357	1,316	1,32
Positive control	0,044	0,041	0,044	0,042	0,04	0,041
Test product	1,213	1,194	1,209	1,424	1,414	1,342

#### 2- Mean, Std Dev and % CV of OD value

	Contact time:24 hours		
	Mean OD value	Standard deviation	CV%
Vehicol control	1,314	0,062	4,718
Negative control	1,263	0,087	6,888
Positive control	0,042	0,002	4,762
Test product	1,299	0,107	8,237

#### 3-Counting of % viability

Vehicol control	Negative Control	Positive Control	Test Product
OD corresponding to 100% viability	Viability %	Viability %	Viability %
1,314	96	3	99

#### 4-Acceptability criteria

	Vehicol control			
mean OD : $\geq 0,3$	VALID			
	Negative control $\geq 70\%$		Positive control $< 70\%$	
Viability	VALID		VALID	
	Vehicol control	Negative control	Positive control	Test product
CV between replicates $< 18\%$	VALID	VALID	VALID	VALID

#### 5- Interpretation of results

Reduction of Viability % measured after contact time	Result
$> 30\%$ after contact time	Cytotoxic
$\leq 30\%$ after contact time	Not Cytotoxic

#### 6- Results

	Reduction Viability %	
Test product	1	NOT CYTOTOXIC

Technician signature 

Date 03/10/18

Supervisor signature 

Date 05/10/18

	<b>Medical Device Testing</b>	<b>Test Facility Eurofins Biolab S.r.l.</b>	Page: <span style="float: right;">1 of 1</span>
---	-----------------------------------	---	---

<b>ADDENDUM N.2 TO THE REPORT: TEST FOR IN VITRO CYTOTOXICITY – DIRECT CONTACT</b>													
<b>REFERENCE/GUIDELINE:</b>	- ISO 10993-5:2009 - Biological evaluation of medical devices Part 5: Tests for in vitro cytotoxicity												
<b>CELL LINE</b>	Mammal fibroblasts BALB/3T3 clone A31 (ATCC®; CCL163™) Source: ATCC.												
<b>MATERIALS</b>	Foetal Bovine Serum (FBS), Trypsin-EDTA, Neutral Red dye, SDS, Ethanol (Sigma-Aldrich)												
	Dulbecco's Modification of Eagle's Medium (DMEM), Dulbecco's Phosphate buffer solution (DPBS), Penicillin/Streptomycin solution (Lonza)												
	Acetic Acid												
	Whatman Inert filter paper disc - diam. 6mm /												
	Water for Injection												
<b>EQUIPMENT</b>	Laminar flow hood, CO <sub>2</sub> incubator, Microplate reader Mod EL800, Chronometer, Common laboratory equipment, Water, Inverted Microscope Diavert, Orbital shaker, Refrigerator												
<b>EXPERIMENTAL DESIGN</b>													
The experimental design included two 12-well plate containing a subconfluent cell monolayer subdivided in the following groups:													
	<table border="1" style="width: 100%; text-align: center;"> <tr> <td><i>Vehicle</i></td> <td><i>Negative control</i></td> <td><i>Positive control</i></td> <td><i>Test sample</i></td> </tr> <tr> <td><i>Vehicle</i></td> <td><i>Negative control</i></td> <td><i>Positive control</i></td> <td><i>Test sample</i></td> </tr> <tr> <td><i>Vehicle</i></td> <td><i>Negative control</i></td> <td><i>Positive control</i></td> <td><i>Test sample</i></td> </tr> </table>	<i>Vehicle</i>	<i>Negative control</i>	<i>Positive control</i>	<i>Test sample</i>	<i>Vehicle</i>	<i>Negative control</i>	<i>Positive control</i>	<i>Test sample</i>	<i>Vehicle</i>	<i>Negative control</i>	<i>Positive control</i>	<i>Test sample</i>
<i>Vehicle</i>	<i>Negative control</i>	<i>Positive control</i>	<i>Test sample</i>										
<i>Vehicle</i>	<i>Negative control</i>	<i>Positive control</i>	<i>Test sample</i>										
<i>Vehicle</i>	<i>Negative control</i>	<i>Positive control</i>	<i>Test sample</i>										
<b>VEHICLE</b>	Supplemented culture medium (without test sample).												
<b>TEST SAMPLE</b>	See the Report for details. Prior to the testing the sample was stored at room temperature.												
<b>NEGATIVE CONTROL</b>	The negative control was represented by 50 µl of DPBS.												
<b>POSITIVE CONTROL</b>	The positive control was represented by 50 µl of 5% SDS.												
<p><b>TREATMENT:</b> Verified that a subconfluent monolayer was present, supplemented culture medium was replaced and the test sample was added through deposition. The plates were incubated in a thermostat at 37°C ±1°C in a 5% CO<sub>2</sub> atmosphere for 24 hours. This procedure was repeated for positive and negative controls.</p>													
<p><b>QUALITATIVE EVALUATION (GRADE OF CYTOTOXICITY):</b> After 24 hours the plates were observed under an inverted microscope and biological reactions were evaluated following a 0 to 4 scale according to ISO10993-5:2009.</p>													
<p><b>QUANTITATIVE EVALUATION (OPTICAL DENSITY):</b> After microscopic observation the supernatants were removed and cells were treated with Neutral Red Medium for 3 hours at 37°C ±1°C in 5% CO<sub>2</sub> atmosphere. Subsequently, the Neutral Red medium was removed and each well was rinsed with DPBS. The plates were totally made dry reversing the plates, then Desorb Solution was added and the plates were incubated for 10 minutes at room temperature with gentle agitation to form an homogeneous solution. Optical density was measured at 540nm by Gen5 software (Biotek) using microtiter plate reader.</p> $\% \text{ of cell viability} = \frac{\text{OD test sample}}{\text{OD vehicle}} \cdot 100$													
<b>ACCEPTABILITY CRITERIA</b>	QUALITATIVE EVALUATION	Negative control ≤ 1; Positive control ≥ 3											
	QUANTITATIVE EVALUATION	The OD mean of the vehicle must be ≥ 0,3. The positive control % cellular viability must be < 70%. The negative control % cellular viability must be ≥ 70%. Coefficient of variation of each group must be ≤ 18%.											
<b>INTERPRETATION OF RESULTS</b>	The achievement of a numerical grade greater than 2 is considered a cytotoxic effect. A cellular viability reduction more than 30% is considered a cytotoxic effect.												
<b>QUALITY CRITERIA</b>	Satisfied												

This certificate may not be reproduced even in part unless expressly approved in writing by Eurofins Biolab. The test results relate only to the items tested. Sampling, except specific indication on test report, is always intended to be made by the Manufacturer.  
Characterization of the test sample is under the Sponsor's responsibility.

### Eurofins Biolab S.r.l.

Società con Socio unico sottoposta  
a direzione e coordinamento della società  
Eurofins Pharma Services Italia Holding  
parte di Eurofins Scientific Group  
[InfoFarma@eurofins.com](mailto:InfoFarma@eurofins.com)  
[www.eurofins.it](http://www.eurofins.it) [www.biolab.it](http://www.biolab.it)

**Via Bruno Buozzi, 2**  
**20090 Vimodrone (MI)**  
**Tel. + 39-022507151**  
**Fax + 39-0225071599**  
[certificata@pec.biolab.it](mailto:certificata@pec.biolab.it)

C.SOC. € 100.000 i.v.  
P. IVA 00762140960  
C.F. 03765750157  
REA MI 966696  
D-U-N-S 429117112  
CIT005

Reviewed and electronically signed for Technical Supervisor Approval by  
Robert Narloch, Employee  
for Eurofins Biolab Srl, on 05-Oct-2018 17:30:55 UTC+02:00